

# c1vac to XGS600 (successful communication)

Read request command sent from c1vac to XGS600 on pin#3 (via cross-wired cable) on CH1 and pressure reading data returned via pin#2 on CH2

