

E mirror selection

| SN # | Measurement [div] | Measurement [arcsec] | Note |
|---------|----------------------|-------------------------|------|
| E1 | -0.8 | -24. | good |
| E2 | -0.8 | -24. | good |
| E3 | -0.2 | -7. | good |
| E4 | -0.5 | -15. | good |
| E5 | 0.8 | 24. | good |
| E6 | -1.0 | -30. | good |
| E7 | -0.2 | -6. | good |
| E8 | -0.8 | -24. | good |
| E9 | -1.0 | -30. | good |
| E10 | 0.0 | 0. | good |
| E11 | -1.0 | -30. | good |
| E12 | -0.3 | -9. | good |
| E13 | -0.8 | -24. | good |
| E14 | -1.0 | -30. | good |
| E15 | -1.2 | -36. | |
| E16 | -0.7 | -21. | good |
| E17 | -0.8 | -24. | good |
| E18 | -1.0 | -30. | good |

| Prism Mirror E | |
|----------------|--------------|
| SN | Location |
| E 1 | OMC(002) SM2 |
| E 2 | OMC(002) SM3 |
| E 3 | OMC(002) BS1 |
| E 4 | OMC(001) SM2 |
| E 5 | OMC(002) SM1 |
| E 6 | |
| E 7 | OMC(003) BS1 |
| E 8 | OMC(003) SM1 |
| E 9 | |
| E 10 | OMC(001) BS1 |
| E 11 | |
| E 12 | OMC(001) SM1 |
| E 13 | OMC(003) SM2 |
| E 14 | |
| E 15 | |
| E 16 | OMC(001) SM3 |
| E 17 | OMC(003) SM3 |
| E 18 | |

Table 7: Perpendicularity measurement for the Mirror E Table 24: List and locations for Mirror E