



<http://www.thorlabs.com>

Thorlabs GmbH
Hans-Boeckler-Str. 6
85221 Dachau
Germany
Tel.: +49 8131 5956.0
Fax.: +49 8131 5956.99

Certificate of Calibration

No 103561542

X New Unit
Returned Unit

Manufacturer Thorlabs GmbH
Model Number S322C
Serial Number 81244
Date of Calibration 22-Dec-10

Calibration Results

Sensitivity 0,1675 mV/W +/- 3%
Wavelength 1064 nm

Calibration Standards

NO	MANUFACTURER	MODEL	SERIAL NO	CERTIFICATION	LAST CAL
1	Thorlabs	S310C	080958	45403-PTB-09	16-Feb-09
2	Laserpoint	A-40-D25	080752	NIST 814856	03-Sep-08
3	Laserpoint	W-1500	081053	43191-PTB-08	07-Nov-08

Thorlabs GmbH does hereby certify that the above mentioned equipment has been calibrated in accordance with our quality management system.

Our Quality Management System is certified according to DIN EN ISO 9001.

The measurement equipment used for calibration is traceable to national standards of the 'EUROMET' members (NPL, PTB, BNM etc.), the US 'NIST' or other national metrological institutions.

Measurements which cannot be traced to national standards can be traced to natural constants, other accepted standards or relational measurements.

Additional documentation concerning traceability of the measurement equipment is available and can be examined upon request.

The certificate of calibration may only be forwarded in complete form without any changes.

The recommended calibration interval is 12 months. The calibration period of this instrument / system begins on the date of receipt by the customer.

Calibrated by Imanakov

Imanakov Sergej
Thorlabs GmbH

Date Received _____

Calibrated Due _____