

---

**HEALTH AND SAFETY SURVEY  
CALIFORNIA INSTITUTE OF TECHNOLOGY  
ENVIRONMENT, HEALTH AND SAFETY SERVICES**

---

**This survey is conducted to assist you in maintaining a safe and healthful workplace for all employees. Please email your EHS representative within 30 days when the items have been resolved and if you need any assistance in addressing any items noted in this survey.**

**Laboratory**                      LIGO Lab                      **Inspection Date**      3/3/2011  
**Principal Investigator**      Lazzarini, Albert              **EHS Staff Member**    Martinez, Larry  
**Safety Officer/Lab Rep**      King, Peter                      **Comments**

<b>Building</b>	<b>Room</b>	<b>Observation</b>	<b>Comments</b>	<b>Resolved</b>	<b>Resolution</b>	<b>Date</b>
Central Engineering Services	40 Meter	Miscellaneous	Replace household extension cord to fan	<input type="checkbox"/>		
Central Engineering Services	40 Meter	Unlabeled Containers	Nitrogen Cylinder	<input type="checkbox"/>		
Central Engineering Services	40 Meter	Permanent Extension Cords	Extension cords are "Daisy Chained"	<input type="checkbox"/>		
Central Engineering Services	40 Meter	Equipment not Seismically Secured	Instrument Racks	<input type="checkbox"/>		
Central Engineering Services	40 Meter	Obstructed Electrical Panel	Blocked by Work Station	<input type="checkbox"/>		